



Hydrometer Temperature Conversion

| Temperature of sample (°F) | Reading Correction | Temperature of sample (°F) | Reading Correction |
|----------------------------|--------------------|----------------------------|--------------------|
| 50° | -.0005 | 150° | +.019 |
| 60° | .000 | 155° | +.020 |
| 70° | +.001 | 160° | +.022 |
| 77° | +.002 | 165° | +.024 |
| 84° | +.003 | 170° | +.025 |
| 95° | +.005 | 175° | +.027 |
| 105° | +.007 | 180° | +.029 |
| 110° | +.008 | 185° | +.031 |
| 115° | +.009 | 190° | +.032 |
| 120° | +.010 | 195° | +.034 |
| 125° | +.012 | 200° | +.036 |
| 130° | +.013 | 205° | +.038 |
| 135° | +.014 | 210° | +.039 |
| 140° | +.016 | | |
| 145° | +.017 | | |
| 150° | +.019 | | |